

UKCA Declaration of Conformity.

(This UKCA Declaration of Conformity is issued under the sole responsibility of the manufacturer)

Manufacturer: WSR Medical Solutions,
Unit 2 Hargreaves Mill
Hargreaves Street
Haslingden
Lancashire
BB4 5RQ
United Kingdom

Authorised Representative: N/A

“The device(s) listed below and on the attachment fulfils the requirement of the Regulations of MDD 93/42/EEC which apply to the device.”

Device name.	Basic UDI-DI	Product code	Purpose of device	Risk class.	SRN
Lateral Foot Imaging Step	5057326001445	LFIS	Patient Positioning	1	
Lateral Foot Imaging Step Deluxe	5057326002756	LFISD	Patient Positioning	1	
Double Step	5057326001452	S105	Patient Positioning	1	
Single Step	5057326001469	S103	Patient Positioning	1	

RELEVANT HARMONISED STANDARDS: EN ISO 14971:2012, EN ISO 15223-1:2016; **BS EN 14183:2003** – *Step stools – Functional and safety requirements, performance testing, and dimensional criteria.*

GMDN: 58878

RELEVANT COMMON SPECIFICATIONS:

ISSUED BY:

SIGN



NAME: Paul Dixon,
DATE:...AUGUST 2024